TWL Calibration Data Summary

Serial Number: 0301362

Full-Scale Flow Rate:	80 (yields 20mA output)
Zero Flow Rate:	(yields 4mA output)
Bridge Resistance:	350.5 ohms
Full-Scale Bridge Output:	1.981 mV/V
Calibration Resistor (R7):	47198 ohms
Calibration Voltage:	0.9337 volts
Calibration Flow Rate:	77.3
Calibration Current:	19.46 mA
Zero Switch Setting:	0
TB3 Jumper Position:	+ (POSITIVE)
Span Switch Setting - Undamped:	0
Span Switch Setting - Damped:	8
Additional Information:	