

RIA Meeting Notes

Subject: STATUS

Date: 4/26/05

Attendees: SUSHIL FRANK AL PETER BRIAN

Action Items:

57 MHZ

- WALCO COMING IN THURSDAY TO LOOK AT 57 MHZ PARTS

- SERVICE REQUEST FOR 1000 HOURS IN WORKS FOR DESIGNER

12 MHZ INJECTOR CAVITY

- NEW FLANGE MAY BE REQUIRED
- LATER DESIGN TO HAVE SQUARE MODULATED VANE ONLY NO STRONG BACK
- REDO DESIGN TO REUSE ALL CURRENT EQUIPMENT
 - REPLACE VANE TIP ONLY
 - 500 MICRON LOCATIONAL TOLERANCE
 - REMACHINE EXISTING ALUM SECTION TO BETTER FLATNESS TOLERANCE MAY NOT BE POSSIBLE
 - MEET WITH MARK KEDZIE FOR QUESTIONS ON OLD DESIGN ASSEMBLY
 - WOODALLOY - MATERIAL?????
 -