

Summary of FY2006 Operational Statistics

Run Identifier	Scheduled User Hours	Storage Ring Availability [hours]	Storage Ring Availability [%]	X-ray Availability [hours]	X-ray Availability [%]	Injector Availability	Faults	Mean Time to Recovery [hours]	Average Fill Duration without a Fault [hours]	Faults Per Day of Delivered Beam	Delivered Integrated Current [A-hr]	Average Current [ma]
Run 05-3	1593.0	1543.6	96.9%	1543.6	96.9%	95.3%	22	2.25	70.2	0.02	154.9	100.3
Run 06-1	1655.0	1627.6	98.3%	1627.6	98.3%	98.1%	13	2.11	125.2	0.19	163.0	100.1
Run 06-2	1752.0	1704.3	97.3%	1704.3	97.3%	98.4%	22	2.17	74.1	0.31	171.1	100.4
Fiscal Year	4999.9	4875.5	97.5%	4875.5	97.5%	97.3%	57	2.18	85.5	0.28	489.0	100.3

Updated

1/23/2008

Data through end of Run 2006-2