

Summary of FY2001 Operational Statistics

Run Identifier	Scheduled User Hours	Storage Ring Availability [hours]	Storage Ring Availability [%]	X-ray Availability [hours]	X-ray Availability [%]	Faults	Average Fill Duration without aFault [hours]	Faults Per Day of Delivered Beam	Delivered Integrated Current [A-hr]	Average Current [ma]
Run 00-4	1000.1	964.0	96.4%	953.0	95.3%	45	21.2	1.13	79.2	83.1
Run 01-1	1528.2	1487.7	97.4%	1475.2	96.5%	52	28.4	0.85	124.9	84.6
Run 01-2	1240.0	1195.4	96.4%	1184.3	95.5%	35	33.8	0.71	101.5	85.7
Run 01-3	1232.0	1187.6	96.4%	1176.2	95.5%	56	21.0	1.14	104.2	88.6
Fiscal Year	5000.3	4834.8	96.7%	4788.8	95.8%	188	25.5	0.94	409.7	85.5

Updated 10/15/01

Data through End of Fiscal Year