

**Summary of FY2004 Operational Statistics**

Run Identifier	Scheduled User Hours	Storage Ring Availability [hours]	Storage Ring Availability [%]	X-ray Availability [hours]	X-ray Availability [%]	Faults	Average Fill Duration without aFault [hours]	Faults Per Day of Delivered Beam	Delivered Integrated Current [A-hr]	Average Current [ma]
Run 03-3	1666.0	1622.7	97.4%	1622.4	97.4%	29	55.9	0.43	136.2	84.0
Run 04-1	1647.0	1610.0	97.8%	1610.0	97.8%	28	57.5	0.42	159.8	99.2
Run 04-2	1912.0	1880.3	98.3%	1880.1	98.3%	32	58.8	0.41	186.8	99.3
Fiscal Year	5225.0	5112.9	97.9%	5112.5	97.8%	89	57.4	0.42	482.7	94.4

Updated 8/30/2004

Data through end of Run 2004-2